

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081119\
 Data File : VU033752.D
 Acq On : 11 Aug 2019 15:12
 Operator : JC/SP
 Sample : K4289-03RE
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F2B16RE

Quant Time: Aug 12 06:40:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Aug 12 04:45:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	269438	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	273568	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	123648	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	37028	18.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	37.34%#
7) Chloroethane-d5	1.67	69	42543	21.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	42.80%#
11) 1,1-Dichloroethene-d2	2.26	63	52075	15.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	31.22%#
21) 2-Butanone-d5	4.15	46	154742	125.85	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	125.85%
24) Chloroform-d	4.62	84	123502	41.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.82%
26) 1,2-Dichloroethane-d4	5.29	65	82107	44.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.12%
32) Benzene-d6	5.32	84	192164	29.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	59.52%#
36) 1,2-Dichloropropane-d6	6.31	67	78113	39.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	78.86%
41) Toluene-d8	7.55	98	147366	24.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	48.54%#
43) trans-1,3-Dichloropropene-	7.83	79	29342	29.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.66%#
47) 2-Hexanone-d5	8.29	63	110554	107.73	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.73%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	158509	54.08	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.16%
64) 1,2-Dichlorobenzene-d4	11.84	152	90478	43.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.96%

Target Compounds

					Ovalue
5) Vinyl chloride	1.40	62	10633	4.688 ug/L	93
13) Acetone	2.31	43	10251	8.292 ug/L	95
20) cis-1,2-Dichloroethene	4.20	96	6520	3.366 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081119\
Data File : VU033752.D
Acq On : 11 Aug 2019 15:12
Operator : JC/SP
Sample : K4289-03RE
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F2B16RE

Quant Time: Aug 12 06:40:41 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080919WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Aug 12 04:45:27 2019
Response via : Initial Calibration

