


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081119\
 Data File : VU033740.D
 Acq On : 11 Aug 2019 10:18
 Operator : JC/SP
 Sample : BFB53
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB53

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080919WMA.M
 Title : VOC Analysis
 Last Update : Mon Aug 12 18:02:58 2019



AutoFind: Scans 2861, 2862, 2863; Background Corrected with Scan 2850

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	7319	PASS
75	95	30	80	51.0	20386	PASS
95	95	100	100	100.0	40005	PASS
96	95	5	9	6.8	2715	PASS
173	174	0.00	2	0.5	188	PASS
174	95	50	120	87.0	34808	PASS
175	174	5	9	7.8	2731	PASS
176	174	95	101	95.5	33256	PASS
177	176	5	9	6.2	2067	PASS

m/z	Abundance
35.90	359.0
37.00	2027.0
38.00	1613.0
39.00	990.0
39.90	204.0
43.90	562.0
45.00	446.0
46.90	265.0
47.90	201.0
49.00	1412.0
50.00	7075.0
51.00	2448.0
52.00	127.0
55.10	114.0
55.90	593.0
57.00	1056.0
60.00	344.0
61.00	1920.0
62.00	2068.0
63.00	1502.0
64.00	252.0
64.90	301.0
67.00	131.0
68.00	3836.0
69.00	4179.0
70.00	267.0
72.10	248.0
73.00	1902.0
74.00	6838.0
75.00	20320.0
76.00	1636.0
77.00	268.0
77.90	141.0
78.90	1486.0
80.00	368.0
80.90	1477.0
81.80	167.0
87.00	1857.0
88.00	1738.0
90.90	175.0
92.00	1050.0
92.90	1847.0
94.00	4954.0
95.00	39544.0
96.10	2711.0
103.90	153.0
105.80	187.0
114.80	104.0
115.80	128.0
116.90	201.0
117.90	226.0
118.90	292.0
127.90	193.0
129.80	214.0
140.90	346.0
142.90	460.0
144.90	120.0
148.00	137.0
172.00	462.0
174.00	33560.0
175.00	2624.0
176.00	31888.0
176.90	2127.0

Instrument :
MSVOA_U
ClientSampleId :
BFB53

m/z	Abundance
35.90	376.0
37.00	2152.0
38.00	1742.0
39.10	904.0
39.90	198.0
44.00	595.0
45.00	449.0
47.00	361.0
47.90	202.0
49.00	1557.0
50.00	7583.0
51.00	2520.0
52.00	127.0
55.00	136.0
56.00	662.0
57.00	1184.0
59.90	357.0
61.00	2039.0
62.00	2110.0
63.00	1614.0
64.00	266.0
65.00	309.0
67.00	130.0
68.00	3764.0
69.00	4214.0
70.00	336.0
72.10	249.0
73.00	1933.0
74.00	7048.0
75.00	20976.0
76.00	1563.0
77.00	238.0
77.90	174.0
78.90	1525.0
80.00	382.0
81.00	1484.0
81.90	221.0
87.00	1845.0
88.00	1869.0
90.90	205.0
92.00	1098.0
93.00	1871.0
94.00	5054.0
95.00	41328.0
96.10	2804.0
103.90	152.0
105.80	168.0
114.80	121.0
115.80	121.0
116.90	206.0
117.80	174.0
118.90	299.0
127.80	210.0
129.80	220.0
140.90	356.0
142.90	506.0
144.90	148.0
172.00	568.0
174.00	35824.0
175.00	2833.0
176.00	34392.0
176.90	2073.0

Instrument :
MSVOA_U
ClientSampleId :
BFB53

Scan 2850 (10.254 min): VU033740.D
BFB53

m/z	Abundance
40.00	150.0
43.90	259.0

Instrument :
MSVOA_U
ClientSampleId :
BFB53

m/z	Abundance
35.90	294.0
37.00	1911.0
38.00	1832.0
39.00	734.0
39.90	181.0
44.00	646.0
45.00	405.0
47.00	429.0
48.00	190.0
49.00	1604.0
50.00	7299.0
51.00	2382.0
55.00	142.0
56.00	615.0
57.00	1200.0
59.90	321.0
61.00	2000.0
62.00	1902.0
63.00	1595.0
64.00	270.0
64.90	332.0
67.10	119.0
68.00	3473.0
69.00	3974.0
70.00	380.0
72.00	225.0
73.00	1836.0
74.00	6643.0
75.00	19864.0
76.00	1365.0
77.10	192.0
77.90	188.0
78.90	1390.0
79.90	350.0
81.00	1481.0
81.90	247.0
86.90	1704.0
87.90	1799.0
90.90	212.0
92.00	1019.0
93.00	1797.0
94.00	4601.0
95.00	39144.0
96.00	2632.0
103.90	150.0
105.80	154.0
114.80	112.0
116.90	215.0
117.80	116.0
118.90	265.0
127.90	181.0
129.80	177.0
134.90	114.0
141.00	378.0
142.80	516.0
144.90	181.0
172.00	623.0
173.00	564.0
174.00	35040.0
175.00	2737.0
176.00	33488.0
176.90	2001.0

Instrument :
MSVOA_U
ClientSampleId :
BFB53