

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
 Data File : VU050247.D
 Acq On : 11 Aug 2022 23:18
 Operator : SY/MD
 Sample : N4160-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGBW6

Quant Time: Aug 12 05:01:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 04:56:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	96217	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	96639	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	36396	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.598	65	39782	4.968	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.400%
7) Chloroethane-d5	1.909	69	32454	5.024	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.400%
11) 1,1-Dichloroethene-d2	2.565	65	14167	4.527	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.600%
20) 2-Butanone-d5	4.633	46	99506	50.960	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.920%
24) Chloroform-d	5.064	84	69050	4.636	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.703	65	36345	4.753	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
32) Benzene-d6	5.729	84	135067	4.479	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
36) 1,2-Dichloropropane-d6	6.691	67	45983	4.630	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.600%
41) Toluene-d8	7.900	98	104623	3.928	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.600%
43) trans-1,3-Dichloroprop...	8.179	79	14407	4.048	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.000%
46) 2-Hexanone-d5	8.636	63	55059	43.629	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.260%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	39158	4.490	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	33815	4.742	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.800%
Target Compounds						Qvalue
3) Chloromethane	1.520	50	1463	0.127	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
Data File : VU050247.D
Acq On : 11 Aug 2022 23:18
Operator : SY/MD
Sample : N4160-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGBW6

Quant Time: Aug 12 05:01:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 04:56:24 2022
Response via : Initial Calibration

