

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
 Data File : VU050258.D
 Acq On : 12 Aug 2022 03:48
 Operator : SY/MD
 Sample : N4161-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGBZ2

Quant Time: Aug 12 05:04:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 04:56:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	91934	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	94891	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	35316	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	39090	5.109	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.200%
7) Chloroethane-d5	1.913	69	32502	5.266	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.400%
11) 1,1-Dichloroethene-d2	2.565	65	13855	4.634	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.600%
20) 2-Butanone-d5	4.639	46	101093	54.184	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.360%
24) Chloroform-d	5.064	84	68840	4.837	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.703	65	37320	5.108	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
32) Benzene-d6	5.729	84	131151	4.429	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.690	67	45699	4.686	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.800%
41) Toluene-d8	7.899	98	102079	3.903	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.000%
43) trans-1,3-Dichloroprop...	8.182	79	13612	3.895	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.000%
46) 2-Hexanone-d5	8.636	63	49374	39.845	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.680%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	40651	4.747	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	33439	4.833	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
Data File : VU050258.D
Acq On : 12 Aug 2022 03:48
Operator : SY/MD
Sample : N4161-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGBZ2

Quant Time: Aug 12 05:04:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 04:56:24 2022
Response via : Initial Calibration

