

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
 Data File : VU050270.D
 Acq On : 12 Aug 2022 10:22
 Operator : SY/MD
 Sample : N4096-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJM5

Quant Time: Aug 12 23:13:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 23:11:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	77558	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	81238	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	31084	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	38870	6.022	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	120.400%	
7) Chloroethane-d5	1.912	69	32480	6.237	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	124.800%	
11) 1,1-Dichloroethene-d2	2.565	65	14262	5.654	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	113.000%	
20) 2-Butanone-d5	4.642	46	94961	60.332	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	120.660%	
24) Chloroform-d	5.063	84	69818	5.816	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	116.400%	
26) 1,2-Dichloroethane-d4	5.703	65	36908	5.988	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	119.800%	
32) Benzene-d6	5.729	84	132388	5.222	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.400%	
36) 1,2-Dichloropropane-d6	6.693	67	45813	5.488	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	109.800%	
41) Toluene-d8	7.899	98	101257	4.522	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.400%	
43) trans-1,3-Dichloroprop...	8.179	79	14683	4.908	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	98.200%	
46) 2-Hexanone-d5	8.639	63	43681	41.175	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	82.340%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	40580	5.535	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	110.800%	
66) 1,2-Dichlorobenzene-d4	12.195	152	34243	5.623	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	112.400%	
Target Compounds						Qvalue
3) Chloromethane	1.517	50	1336	0.144	ug/L	98
13) Acetone	2.658	43	5357	6.111	ug/L	72
34) Trichloroethene	6.546	95	7039	1.023	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
Data File : VU050270.D
Acq On : 12 Aug 2022 10:22
Operator : SY/MD
Sample : N4096-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBJM5

Quant Time: Aug 12 23:13:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 23:11:34 2022
Response via : Initial Calibration

