

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
 Data File : VU050295.D
 Acq On : 12 Aug 2022 21:18
 Operator : SY/MD
 Sample : N4072-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUX4

Quant Time: Aug 13 01:24:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 13 01:20:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	97253	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	96763	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	38701	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	34293	4.237	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.800%
7) Chloroethane-d5	1.909	69	29815	4.566	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.400%
11) 1,1-Dichloroethene-d2	2.562	65	13207	4.175	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.600%
20) 2-Butanone-d5	4.645	46	109344	55.401	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.800%
24) Chloroform-d	5.063	84	64688	4.297	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
26) 1,2-Dichloroethane-d4	5.703	65	34280	4.435	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.800%
32) Benzene-d6	5.729	84	120339	3.985	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.800%
36) 1,2-Dichloropropane-d6	6.690	67	43436	4.368	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.400%
41) Toluene-d8	7.899	98	91296	3.423	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.400%#
43) trans-1,3-Dichloroprop...	8.182	79	13479	3.783	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.600%
46) 2-Hexanone-d5	8.639	63	67652	53.539	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.080%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	38177	4.372	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.400%
66) 1,2-Dichlorobenzene-d4	12.192	152	33956	4.478	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.600%
Target Compounds						Qvalue
3) Chloromethane	1.517	50	1551	0.133	ug/L	97
13) Acetone	2.668	43	6133	5.580	ug/L	67
14) Carbon disulfide	2.787	76	5134	0.214	ug/L	99
17) Methyl tert-butyl Ether	3.366	73	452072	25.767	ug/L	99
18) trans-1,2-Dichloroethene	3.356	96	2294	0.305	ug/L	93
22) cis-1,2-Dichloroethene	4.668	96	6275	0.751	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 13 01:24:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 13 01:20:27 2022
Response via : Initial Calibration

