

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
 Data File : VU050229.D
 Acq On : 11 Aug 2022 14:40
 Operator : SY/MD
 Sample : N4160-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGBS9

Quant Time: Aug 12 01:32:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 01:30:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	82250	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	83425	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	32455	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	42131	6.155	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 123.200%	
7) Chloroethane-d5	1.912	69	35478	6.424	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 128.400%	
11) 1,1-Dichloroethene-d2	2.565	65	16284	6.087	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 121.800%	
20) 2-Butanone-d5	4.642	46	114797	68.774	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	= 137.540%#	
24) Chloroform-d	5.063	84	74558	5.856	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 117.200%	
26) 1,2-Dichloroethane-d4	5.703	65	40192	6.148	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 123.000%	
32) Benzene-d6	5.729	84	146755	5.637	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 112.800%	
36) 1,2-Dichloropropane-d6	6.694	67	49660	5.793	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	= 115.800%	
41) Toluene-d8	7.899	98	117293	5.101	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 102.000%	
43) trans-1,3-Dichloroprop...	8.182	79	16772	5.459	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	= 109.200%	
46) 2-Hexanone-d5	8.639	63	62719	57.570	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	= 115.140%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	43391	5.764	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 115.200%	
66) 1,2-Dichlorobenzene-d4	12.195	152	37833	5.950	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 119.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
Data File : VU050229.D
Acq On : 11 Aug 2022 14:40
Operator : SY/MD
Sample : N4160-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGBS9

Quant Time: Aug 12 01:32:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 01:30:48 2022
Response via : Initial Calibration

