

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
 Data File : VU050252.D
 Acq On : 12 Aug 2022 01:21
 Operator : SY/MD
 Sample : N4161-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGBY6

Quant Time: Aug 12 05:30:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 12 04:56:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	84133	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	86596	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	31122	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	39899	5.699	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.000%
7) Chloroethane-d5	1.912	69	32072	5.678	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.600%
11) 1,1-Dichloroethene-d2	2.565	65	14079	5.145	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.000%
20) 2-Butanone-d5	4.642	46	105095	61.552	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.100%
24) Chloroform-d	5.063	84	68973	5.296	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.000%
26) 1,2-Dichloroethane-d4	5.703	65	37355	5.586	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.800%
32) Benzene-d6	5.726	84	131769	4.876	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
36) 1,2-Dichloropropane-d6	6.690	67	45535	5.117	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.400%
41) Toluene-d8	7.899	98	103251	4.326	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.600%
43) trans-1,3-Dichloroprop...	8.182	79	13749	4.312	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.200%
46) 2-Hexanone-d5	8.639	63	52528	46.450	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.900%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	40091	5.130	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	33518	5.497	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081122\
Data File : VU050252.D
Acq On : 12 Aug 2022 01:21
Operator : SY/MD
Sample : N4161-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGBY6

Quant Time: Aug 12 05:30:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 12 04:56:24 2022
Response via : Initial Calibration

