

Data File : VU033771.D
Acq On : 12 Aug 2019 15:01
Operator : JC/SP
Sample : K4202-15RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F2B40RE

Quant Time: Aug 12 17:07:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080919WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Aug 12 16:26:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	387886	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	385702	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	170296	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	107185	37.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.08%
7) Chloroethane-d5	1.67	69	96649	33.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	67.54%#
11) 1,1-Dichloroethene-d2	2.26	63	142580	29.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.40%#
21) 2-Butanone-d5	4.16	46	95685	54.05	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	54.05%
24) Chloroform-d	4.62	84	149928	35.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	70.68%
26) 1,2-Dichloroethane-d4	5.29	65	92826	34.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	69.98%#
32) Benzene-d6	5.32	84	302849	33.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	66.54%#
36) 1,2-Dichloropropane-d6	6.31	67	95171	34.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	68.14%#
41) Toluene-d8	7.55	98	252441	29.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	58.96%#
43) trans-1,3-Dichloropropene-	7.83	79	39315	27.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.74%#
47) 2-Hexanone-d5	8.30	63	64266	44.42	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	44.42%#
57) 1,1,2,2-Tetrachloroethane-	10.41	84	125728	30.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	60.86%#
64) 1,2-Dichlorobenzene-d4	11.85	152	92436	32.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	64.50%#

Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	4.20	96	34934	12.528	ug/L	98
34) Trichloroethene	6.17	95	38133	13.464	ug/L	94
46) Tetrachloroethene	8.21	164	21298	9.123	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU033771.D
Acq On : 12 Aug 2019 15:01
Operator : JC/SP
Sample : K4202-15RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F2B40RE

Quant Time: Aug 12 17:07:20 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080919WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Aug 12 16:26:52 2019
Response via : Initial Calibration

