


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU081220\
 Data File : VU039836.D
 Acq On : 12 Aug 2020 09:52
 Operator : SY/MD
 Sample : BFB
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U072920DW.M
 Title : METHOD 524.2 VOLATILES DRINKING WATER
 Last Update : Fri Jul 31 07:12:27 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	7054	PASS
75	95	30	60	55.9	18429	PASS
95	95	100	100	100.0	32957	PASS
96	95	5	9	6.9	2264	PASS
173	174	0.00	2	0.5	131	PASS
174	95	50	100	82.1	27050	PASS
175	174	5	9	7.8	2123	PASS
176	174	95	101	95.5	25824	PASS
177	176	5	9	6.6	1714	PASS