

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081224\
 Data File : VU060405.D
 Acq On : 12 Aug 2024 14:44
 Operator : MD/SY
 Sample : P3540-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B42

Quant Time: Aug 13 02:24:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 13 02:20:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	88130	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	95749	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	40341	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	24399	3.527	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.600%
7) Chloroethane-d5	1.911	69	24991	4.006	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.200%
11) 1,1-Dichloroethene-d2	2.560	65	10504	3.568	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.400%
20) 2-Butanone-d5	4.634	46	76901	52.380	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.760%
24) Chloroform-d	5.055	84	61270	4.473	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
26) 1,2-Dichloroethane-d4	5.695	65	32029	5.042	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.721	84	110977	3.843	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.800%
36) 1,2-Dichloropropane-d6	6.682	67	39351	4.393	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.800%
41) Toluene-d8	7.891	98	88954	3.450	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.000%#
43) trans-1,3-Dichloroprop...	8.181	79	11779	4.068	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.400%
46) 2-Hexanone-d5	8.631	63	58254	45.333	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.660%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	32615	4.438	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	35362	4.758	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081224\
Data File : VU060405.D
Acq On : 12 Aug 2024 14:44
Operator : MD/SY
Sample : P3540-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B42

Quant Time: Aug 13 02:24:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 13 02:20:56 2024
Response via : Initial Calibration

