

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081224\
 Data File : VU060410.D
 Acq On : 12 Aug 2024 17:10
 Operator : MD/SY
 Sample : P3540-03 100X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B41

Quant Time: Aug 13 02:26:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 13 02:20:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	78541	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	83122	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	34768	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	21975	3.564	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.200%
7) Chloroethane-d5	1.908	69	23245	4.181	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.600%
11) 1,1-Dichloroethene-d2	2.560	65	10116	3.855	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.200%
20) 2-Butanone-d5	4.641	46	73818	56.419	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.840%
24) Chloroform-d	5.055	84	57803	4.735	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	5.695	65	30176	5.331	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.600%
32) Benzene-d6	5.721	84	101556	4.051	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.000%
36) 1,2-Dichloropropane-d6	6.686	67	35938	4.622	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.891	98	80476	3.596	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.000%
43) trans-1,3-Dichloroprop...	8.177	79	11073	4.405	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.000%
46) 2-Hexanone-d5	8.631	63	54383	48.749	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.500%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	31039	4.865	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	32615	5.092	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081224\
Data File : VU060410.D
Acq On : 12 Aug 2024 17:10
Operator : MD/SY
Sample : P3540-03 100X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B41

Quant Time: Aug 13 02:26:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 13 02:20:56 2024
Response via : Initial Calibration

