


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081318\
Data File : VU026053.D
Acq On : 13 Aug 2018 07:56
Operator : MD/SY
Sample : BFB45
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB45

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080918WMA.M
Title : VOC Analysis
Last Update : Tue Aug 14 02:18:47 2018



AutoFind: Scans 3022, 3023, 3024; Background Corrected with Scan 3011

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.6	5276	PASS
75	95	30	80	53.4	15139	PASS
95	95	100	100	100.0	28341	PASS
96	95	5	9	7.4	2098	PASS
173	174	0.00	2	1.7	431	PASS
174	95	50	120	88.3	25016	PASS
175	174	5	9	7.9	1967	PASS
176	174	95	101	96.9	24253	PASS
177	176	5	9	6.4	1541	PASS

Scan 3011 (10.271 min): VU026053.D
BFB45

m/z	Abundance
43.90	369.0
206.90	139.0

Instrument :
MSVOA_U
ClientSampleId :
BFB45

m/z	Abundance
36.10	236.0
37.00	1299.0
38.00	1257.0
39.10	556.0
40.00	206.0
44.00	393.0
45.00	221.0
47.10	442.0
47.90	137.0
49.00	1266.0
50.00	5248.0
51.00	1509.0
51.90	101.0
56.10	350.0
57.00	724.0
59.90	297.0
61.00	1387.0
62.00	1511.0
63.00	1258.0
64.00	184.0
65.00	212.0
68.00	2740.0
69.00	2988.0
70.00	268.0
71.90	103.0
73.00	1003.0
74.00	4466.0
75.00	14475.0
76.00	1077.0
76.90	110.0
78.10	115.0
78.90	979.0
79.90	215.0
80.90	1194.0
81.80	235.0
86.90	1183.0
88.00	903.0
92.00	867.0
93.00	1685.0
94.00	3281.0
95.00	27680.0
96.00	1978.0
103.90	227.0
105.80	149.0
115.90	123.0
116.90	234.0
117.90	132.0
118.90	258.0
127.90	148.0
140.90	283.0
142.80	588.0
172.00	362.0
173.00	441.0
174.00	23496.0
175.00	1881.0
176.00	22624.0
177.00	1555.0

Instrument :
MSVOA_U
ClientSampleId :
BFB45

m/z	Abundance
36.10	246.0
37.00	1389.0
38.00	1448.0
39.00	523.0
40.00	207.0
44.00	445.0
45.00	210.0
47.00	476.0
47.80	124.0
49.00	1329.0
50.00	5374.0
51.00	1622.0
56.10	408.0
57.00	730.0
60.00	328.0
61.00	1336.0
62.00	1654.0
63.00	1287.0
63.90	224.0
65.00	181.0
68.00	3025.0
69.00	3382.0
70.00	231.0
72.00	143.0
73.00	1220.0
74.00	4796.0
75.00	15556.0
76.00	1149.0
76.80	165.0
78.10	106.0
78.90	968.0
79.90	192.0
80.90	1276.0
81.80	252.0
86.90	1239.0
87.90	967.0
90.90	152.0
92.00	888.0
93.00	1823.0
94.00	3518.0
95.00	29048.0
96.00	2166.0
103.90	249.0
105.80	162.0
116.00	117.0
116.90	281.0
117.90	133.0
118.90	240.0
127.90	151.0
141.00	264.0
142.80	583.0
172.00	371.0
172.90	478.0
174.00	25808.0
174.90	2025.0
176.00	25120.0
176.90	1581.0
178.00	150.0

Instrument :
MSVOA_U
ClientSampleId :
BFB45

m/z	Abundance
36.10	207.0
37.00	1403.0
38.00	1390.0
39.00	485.0
40.00	165.0
44.10	524.0
45.00	173.0
47.00	464.0
47.80	113.0
49.10	1153.0
50.00	5207.0
51.00	1623.0
56.10	367.0
57.10	733.0
60.00	326.0
61.00	1195.0
62.00	1601.0
63.00	1240.0
63.90	208.0
64.90	220.0
68.00	3050.0
69.00	3398.0
69.90	182.0
72.00	157.0
73.00	1316.0
74.00	4935.0
75.00	15386.0
76.00	1248.0
76.80	216.0
77.80	159.0
78.90	857.0
79.90	219.0
80.90	1179.0
81.80	234.0
86.90	1164.0
87.90	952.0
90.90	212.0
92.10	803.0
93.00	1564.0
94.00	3384.0
95.00	28296.0
96.00	2150.0
103.90	211.0
105.80	144.0
116.00	101.0
116.90	307.0
118.90	193.0
127.90	128.0
141.10	209.0
142.90	455.0
172.00	420.0
172.90	376.0
174.00	25744.0
174.90	1996.0
176.00	25016.0
176.90	1488.0
178.00	153.0

Instrument :
MSVOA_U
ClientSampleId :
BFB45