

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081322\
 Data File : VU050319.D
 Acq On : 13 Aug 2022 15:25
 Operator : SY/MD
 Sample : N4115-02DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5E01DL

Quant Time: Aug 14 01:27:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sun Aug 14 01:24:58 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.250	114	89920	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.417	117	88363	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.812	152	34220	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.597	65	44789	5.985	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	119.800%	
7) Chloroethane-d5		1.909	69	35631	5.902	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	118.000%	
11) 1,1-Dichloroethene-d2		2.565	65	16505	5.643	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	112.800%	
20) 2-Butanone-d5		4.629	46	106141	58.164	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.320%	
24) Chloroform-d		5.063	84	74511	5.353	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%	
26) 1,2-Dichloroethane-d4		5.703	65	39547	5.534	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.600%	
32) Benzene-d6		5.729	84	150732	5.466	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.400%	
36) 1,2-Dichloropropane-d6		6.690	67	50534	5.565	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.400%	
41) Toluene-d8		7.899	98	118513	4.866	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%	
43) trans-1,3-Dichloroprop...		8.179	79	16774	5.155	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.000%	
46) 2-Hexanone-d5		8.635	63	60460	52.395	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.800%	
56) 1,1,2,2-Tetrachloroeth...		10.758	84	42935	5.384	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.600%	
66) 1,2-Dichlorobenzene-d4		12.195	152	37961	5.662	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.200%	
Target Compounds							Qvalue
5) Vinyl chloride		1.604	62	19201	1.988	ug/L	97
22) cis-1,2-Dichloroethene		4.668	96	38416	4.973	ug/L	98
34) Trichloroethene		6.546	95	1427	0.191	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081322\
Data File : VU050319.D
Acq On : 13 Aug 2022 15:25
Operator : SY/MD
Sample : N4115-02DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5E01DL

Quant Time: Aug 14 01:27:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sun Aug 14 01:24:58 2022
Response via : Initial Calibration

