

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081322\
 Data File : VU050326.D
 Acq On : 13 Aug 2022 18:18
 Operator : SY/MD
 Sample : N4115-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5D81

Quant Time: Aug 14 01:28:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sun Aug 14 01:24:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	82005	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	81373	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	31081	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	40184	5.888	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 117.800%		
7) Chloroethane-d5	1.912	69	31337	5.691	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 113.800%		
11) 1,1-Dichloroethene-d2	2.565	65	14168	5.312	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 106.200%		
20) 2-Butanone-d5	4.620	46	96654	58.078	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 116.160%		
24) Chloroform-d	5.063	84	64613	5.090	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 101.800%		
26) 1,2-Dichloroethane-d4	5.703	65	35408	5.433	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 108.600%		
32) Benzene-d6	5.729	84	126208	4.970	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 99.400%		
36) 1,2-Dichloropropane-d6	6.690	67	43466	5.198	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 104.000%		
41) Toluene-d8	7.899	98	97997	4.369	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 87.400%		
43) trans-1,3-Dichloroprop...	8.179	79	14009	4.675	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 93.600%		
46) 2-Hexanone-d5	8.632	63	57523	54.132	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 108.260%		
56) 1,1,2,2-Tetrachloroeth...	10.754	84	39118	5.327	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 106.600%		
66) 1,2-Dichlorobenzene-d4	12.192	152	32493	5.336	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 106.800%		
Target Compounds						
5) Vinyl chloride	1.604	62	79765	9.054	ug/L	100
18) trans-1,2-Dichloroethene	3.353	96	7547	1.189	ug/L	95
22) cis-1,2-Dichloroethene	4.668	96	2472	0.351	ug/L	95
33) Benzene	5.774	78	13580	0.476	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081322\
Data File : VU050326.D
Acq On : 13 Aug 2022 18:18
Operator : SY/MD
Sample : N4115-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5D81

Quant Time: Aug 14 01:28:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sun Aug 14 01:24:58 2022
Response via : Initial Calibration

