

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
 Data File : VU060424.D
 Acq On : 13 Aug 2024 14:27
 Operator : MD/SY
 Sample : P3501-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BW5

Quant Time: Aug 14 01:21:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 14 01:18:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	82306	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	87600	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	37103	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	21562	3.337	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.800%
7) Chloroethane-d5	1.911	69	21853	3.751	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.000%
11) 1,1-Dichloroethene-d2	2.560	65	9428	3.429	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.600%
20) 2-Butanone-d5	4.631	46	70314	51.283	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.560%
24) Chloroform-d	5.055	84	55384	4.329	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
26) 1,2-Dichloroethane-d4	5.695	65	28634	4.827	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
32) Benzene-d6	5.721	84	98089	3.713	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.200%
36) 1,2-Dichloropropane-d6	6.682	67	35644	4.350	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.000%
41) Toluene-d8	7.894	98	75199	3.188	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.800%#
43) trans-1,3-Dichloroprop...	8.177	79	9871	3.726	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.600%
46) 2-Hexanone-d5	8.631	63	48612	41.348	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.700%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	29591	4.401	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	31023	4.539	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
Data File : VU060424.D
Acq On : 13 Aug 2024 14:27
Operator : MD/SY
Sample : P3501-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BW5

Quant Time: Aug 14 01:21:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 14 01:18:59 2024
Response via : Initial Calibration

