

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
 Data File : VU060427.D
 Acq On : 13 Aug 2024 15:41
 Operator : MD/SY
 Sample : P3501-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BW8

Quant Time: Aug 14 01:22:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 14 01:18:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	83531	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	89250	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	37537	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	22378	3.413	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.200%
7) Chloroethane-d5	1.911	69	23272	3.936	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.800%
11) 1,1-Dichloroethene-d2	2.563	65	9711	3.480	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.600%
20) 2-Butanone-d5	4.631	46	79304	56.991	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.980%
24) Chloroform-d	5.055	84	60570	4.665	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	5.695	65	31404	5.216	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.721	84	105506	3.920	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.400%
36) 1,2-Dichloropropane-d6	6.682	67	39286	4.705	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.200%
41) Toluene-d8	7.894	98	85326	3.551	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.000%
43) trans-1,3-Dichloroprop...	8.177	79	10866	4.026	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.600%
46) 2-Hexanone-d5	8.631	63	57304	47.840	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.680%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	33050	4.824	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	34808	5.034	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
Data File : VU060427.D
Acq On : 13 Aug 2024 15:41
Operator : MD/SY
Sample : P3501-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BW8

Quant Time: Aug 14 01:22:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 14 01:18:59 2024
Response via : Initial Calibration

