




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
Data File : VU060437.D
Acq On : 13 Aug 2024 20:10
Operator : MD/SY
Sample : P3583-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0UC1

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/14/2024
Supervised By :Mahesh Dadoda 08/14/2024

Quant Time: Aug 14 01:25:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 14 01:18:59 2024
Response via : Initial Calibration



TIC: VU060437.D\data.ms

(15) Methyl Acetate (T)

2.972min (+ 0.013) 0.43 ug/L

response 900

Ion	Exp%	Act%
43.05	100.00	100.00
74.05	26.60	22.33
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
Data File : VU060437.D
Acq On : 13 Aug 2024 20:10
Operator : MD/SY
Sample : P3583-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0UC1

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/14/2024
Supervised By :Mahesh Dadoda 08/14/2024

Quant Time: Aug 14 01:25:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 14 01:18:59 2024
Response via : Initial Calibration



TIC: VU060437.D\data.ms

(15) Methyl Acetate (T)

2.972min (+ 0.013) 0.91 ug/L m

response 1884

Ion	Exp%	Act%
43.05	100.00	100.00
74.05	26.60	10.67#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
 Data File : VU060437.D
 Acq On : 13 Aug 2024 20:10
 Operator : MD/SY
 Sample : P3583-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0UC1

Manual IntegrationsAPPROVED

Quant Time: Aug 14 01:25:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 14 01:18:59 2024
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 08/14/2024
 Supervised By :Mahesh Dadoda 08/14/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	87844	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	88982	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	38841	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	21752	3.155	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	63.000%		
7) Chloroethane-d5	1.907	69	23578	3.792	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	75.800%		
11) 1,1-Dichloroethene-d2	2.563	65	10426	3.553	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	71.000%		
20) 2-Butanone-d5	4.628	46	81501	55.694	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	111.380%		
24) Chloroform-d	5.055	84	62263	4.560	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	91.200%		
26) 1,2-Dichloroethane-d4	5.695	65	32612	5.151	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	103.000%		
32) Benzene-d6	5.721	84	114644	4.272	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	85.400%		
36) 1,2-Dichloropropane-d6	6.682	67	39152	4.704	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	94.000%		
41) Toluene-d8	7.891	98	87502	3.652	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	73.000%		
43) trans-1,3-Dichloroprop...	8.177	79	11431	4.248	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	85.000%		
46) 2-Hexanone-d5	8.627	63	57994	48.562	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	97.120%		
56) 1,1,2,2-Tetrachloroeth...	10.749	84	33607	4.920	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	98.400%		
66) 1,2-Dichlorobenzene-d4	12.187	152	37279	5.210	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	104.200%		
Target Compounds						
						Qvalue
13) Acetone	2.640	43	74163	74.216	ug/L	97
15) Methyl Acetate	2.972	43	1884m	0.910	ug/L	
21) 2-Butanone	4.721	43	17595	11.814	ug/L	97
33) Benzene	5.775	78	5769	0.225	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
Data File : VU060437.D
Acq On : 13 Aug 2024 20:10
Operator : MD/SY
Sample : P3583-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0UC1

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/14/2024
Supervised By :Mahesh Dadoda 08/14/2024

Quant Time: Aug 14 01:25:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 14 01:18:59 2024
Response via : Initial Calibration

