

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
 Data File : VU060421.D
 Acq On : 13 Aug 2024 13:14
 Operator : MD/SY
 Sample : P3501-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BW2

Quant Time: Aug 14 01:21:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 14 01:18:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	82871	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	85920	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	37177	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	20615	3.169	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.400%
7) Chloroethane-d5	1.907	69	22463	3.830	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.600%
11) 1,1-Dichloroethene-d2	2.560	65	10052	3.631	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.600%
20) 2-Butanone-d5	4.634	46	57598	41.722	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.440%
24) Chloroform-d	5.055	84	56966	4.423	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.698	65	27993	4.687	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.800%
32) Benzene-d6	5.721	84	101704	3.925	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.600%
36) 1,2-Dichloropropane-d6	6.685	67	35638	4.434	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.600%
41) Toluene-d8	7.891	98	79443	3.434	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.600%#
43) trans-1,3-Dichloroprop...	8.177	79	9902	3.811	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.200%
46) 2-Hexanone-d5	8.631	63	52694	45.697	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.400%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	31088	4.714	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	31920	4.661	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.200%
Target Compounds						
					Qvalue	
47) Tetrachloroethene	8.547	164	16229	3.204	ug/L	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081324\
Data File : VU060421.D
Acq On : 13 Aug 2024 13:14
Operator : MD/SY
Sample : P3501-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BW2

Quant Time: Aug 14 01:21:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 14 01:18:59 2024
Response via : Initial Calibration





#47

Tetrachloroethene

Concen: 3.204 ug/L

RT: 8.547 min Scan# 21

Delta R.T. 0.000 min

Lab File: VU060421.D


Acq: 13 Aug 2024 13:14

Instrument :

MSVOA_U

ClientSampleId :

C0BW2



Tgt Ion:164 Resp: 16229

Ion Ratio Lower Upper

164 100

129 90.4 69.0 128.1

131 84.4 62.6 116.3

166 114.9 89.2 165.8

