

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081325\
 Data File : VU063557.D
 Acq On : 13 Aug 2025 11:48
 Operator : MD/SY
 Sample : VSTD0.564
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5064

Quant Time: Aug 13 22:39:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081325.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 13 22:38:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	15594	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	20962	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	7345	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.590	65	4615	0.427	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	2776	0.471	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	2878	0.533	ug/L	0.00
8) Toluene-d8	7.887	98	9292	0.507	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	2182	0.522	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	8173	0.476	ug/L	100
6) Trichloroethene	6.544	95	5838	0.459	ug/L	100
9) 1,2-Dibromoethane	8.924	107	3568	0.505	ug/L	100
12) 1,2,3-Trichloropropane	10.814	75	3122	0.487	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.990	75	512	0.456	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081325\
Data File : VU063557.D
Acq On : 13 Aug 2025 11:48
Operator : MD/SY
Sample : VSTD0.564
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5064

Quant Time: Aug 13 22:39:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081325.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 13 22:38:52 2025
Response via : Initial Calibration

