

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081325\
 Data File : VU063558.D
 Acq On : 13 Aug 2025 12:22
 Operator : MD/SY
 Sample : VSTD00165
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD001065

Quant Time: Aug 13 22:39:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081325.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 13 22:38:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	15250	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	20674	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	7173	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.590	65	8906	0.843	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.685	65	5442	0.945	ug/L	-0.02
7) 1,2-Dichloropropane-d6	6.687	67	5608	1.052	ug/L	0.00
8) Toluene-d8	7.887	98	18283	1.012	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	4246	1.031	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.590	62	15596	0.929	ug/L	92
6) Trichloroethene	6.544	95	11563	0.921	ug/L	99
9) 1,2-Dibromoethane	8.924	107	6971	1.001	ug/L #	99
12) 1,2,3-Trichloropropane	10.814	75	6348	1.013	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.990	75	1019	0.930	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081325\
Data File : VU063558.D
Acq On : 13 Aug 2025 12:22
Operator : MD/SY
Sample : VSTD00165
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD001065

Quant Time: Aug 13 22:39:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081325.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 13 22:38:52 2025
Response via : Initial Calibration

