

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081519\
 Data File : VU033803.D
 Acq On : 15 Aug 2019 11:23
 Operator : JC/SP
 Sample : VU0815WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK02

Quant Time: Aug 16 01:18:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR081419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 16 01:09:16 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	173477	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	180322	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.47	152	79835	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	79973	4.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.40%
7) Chloroethane-d5	1.67	69	68957	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.80%
11) 1,1-Dichloroethene-d2	2.26	65	38693	4.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.80%
20) 2-Butanone-d5	4.19	46	205467	54.16	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.32%
24) Chloroform-d	4.62	84	126746	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
26) 1,2-Dichloroethane-d4	5.29	65	72367	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
32) Benzene-d6	5.31	84	239368	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.40%
36) 1,2-Dichloropropane-d6	6.31	67	79682	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.55	98	193313	3.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.00%
43) trans-1,3-Dichloropropene-	7.84	79	29662	4.27	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.40%
46) 2-Hexanone-d5	8.30	63	144473	52.14	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.28%
56) 1,1,2,2-Tetrachloroethane-	10.42	84	70912	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
66) 1,2-Dichlorobenzene-d4	11.85	152	80548	4.98	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081519\
Data File : VU033803.D
Acq On : 15 Aug 2019 11:23
Operator : JC/SP
Sample : VU0815WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK02

Quant Time: Aug 16 01:18:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR081419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 16 01:09:16 2019
Response via : Initial Calibration

