

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081524\
 Data File : VU060461.D
 Acq On : 15 Aug 2024 11:41
 Operator : MD/SY
 Sample : P3555-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCXW4

Quant Time: Aug 16 04:56:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 16 04:55:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	90377	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	95119	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	41270	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	28929	4.078	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.600%
7) Chloroethane-d5	1.907	69	27592	4.313	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.200%
11) 1,1-Dichloroethene-d2	2.560	65	12888	4.269	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.400%
20) 2-Butanone-d5	4.634	46	69941	46.455	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.920%
24) Chloroform-d	5.055	84	65608	4.671	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	5.695	65	31745	4.873	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.400%
32) Benzene-d6	5.721	84	123146	4.293	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.800%
36) 1,2-Dichloropropane-d6	6.682	67	39609	4.451	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.000%
41) Toluene-d8	7.894	98	104843	4.094	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.800%
43) trans-1,3-Dichloroprop...	8.177	79	13203	4.590	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.800%
46) 2-Hexanone-d5	8.631	63	53538	41.939	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.880%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	31748	4.348	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	37323	4.909	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081524\
Data File : VU060461.D
Acq On : 15 Aug 2024 11:41
Operator : MD/SY
Sample : P3555-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCXW4

Quant Time: Aug 16 04:56:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 16 04:55:30 2024
Response via : Initial Calibration

