

Method Path : Z:\voasrv\HPCHEM1\MSVOA_U\Method\
 Method File : SFAMUSIM081621.M
 Title : TRACE VOA SOM01.0
 Last Update : Tue Aug 17 02:55:50 2021
 Response Via : Initial Calibration

Calibration Files

0.5 =VU044428.D 0.05=VU044426.D 0.1 =VU044427.D 1.0 =VU044429.D 2 =VU044430.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.684	0.576	0.692	0.663	0.654	0.654	7.04
3) T	Vinyl chloride	0.533	0.422	0.531	0.490	0.532	0.501	9.60
4) S	1,2-Dichloroetha...	0.722	0.841	0.815	0.694	0.684	0.751	9.57
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.528	0.436	0.491	0.512	0.546	0.503	8.48
7) S	1,2-Dichloroprop...	0.623	0.545	0.647	0.639	0.619	0.615	6.57
8) S	Toluene-d8	1.389	1.382	1.454	1.414	1.395	1.407	2.04
9) T	1,2-Dibromoethane	0.376	0.332	0.368	0.361	0.393	0.366	6.12
10) S	1,1,2,2-Tetrachl...	0.518	0.472	0.559	0.531	0.529	0.522	6.07
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.787	0.660	0.745	0.738	0.801	0.746	7.37
13) T	1,2-Dibromo-3-ch...	0.152	0.112	0.138	0.138	0.156	0.139	12.42

(#) = Out of Range