

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081621\
 Data File : VU044430.D
 Acq On : 16 Aug 2021 13:08
 Operator : SY/MD
 Sample : VSTD00214
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD002014

Quant Time: Aug 17 02:48:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 17 02:46:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	6102	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.364	117	5804	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	2877	0.500	ug/L	-0.01
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.605	65	15964	2.008	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.661	65	16701	1.825	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.639	67	14360	2.013	ug/L	0.00
8) Toluene-d8	7.849	98	32381	1.983	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.682	84	12271	2.027	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.605	62	12975	2.124	ug/L	97
6) Trichloroethene	6.508	95	12676	2.173	ug/L	97
9) 1,2-Dibromoethane	8.869	107	9132	2.149	ug/L #	99
12) 1,2,3-Trichloropropane	10.751	75	9217	2.146	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.927	75	1797	2.243	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081621\
Data File : VU044430.D
Acq On : 16 Aug 2021 13:08
Operator : SY/MD
Sample : VSTD00214
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD002014

Quant Time: Aug 17 02:48:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 17 02:46:12 2021
Response via : Initial Calibration

