

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081621\
 Data File : VU044433.D
 Acq On : 16 Aug 2021 16:12
 Operator : SY/MD
 Sample : M3395-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 X3623

Quant Time: Aug 17 03:03:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 17 02:56:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	6161	0.500	ug/L	0.02
5) Chlorobenzene-d5	9.361	117	5795	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.759	152	2897	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	4293	0.533	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 106.000%			
4) 1,2-Dichloroethane-d4	5.680	65	4563	0.493	ug/L	0.02
Spiked Amount 0.500	Range 70 - 130		Recovery = 98.000%			
7) 1,2-Dichloropropane-d6	6.639	67	3864	0.543	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 108.000%			
8) Toluene-d8	7.845	98	8428	0.517	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 104.000%			
10) 1,1,2,2-Tetrachloroeth...	10.696	84	3287	0.544	ug/L	0.02
Spiked Amount 0.500	Range 65 - 120		Recovery = 108.000%			
Target Compounds						
						Qvalue
3) Vinyl chloride	1.606	62	4905	0.794	ug/L	95
6) Trichloroethene	6.508	95	11559	1.984	ug/L	97
9) 1,2-Dibromoethane	8.866	107	2849	0.672	ug/L	98
12) 1,2,3-Trichloropropane	10.748	75	1058	0.245	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.924	75	219	0.272	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081621\
Data File : VU044433.D
Acq On : 16 Aug 2021 16:12
Operator : SY/MD
Sample : M3395-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
X3623

Quant Time: Aug 17 03:03:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 17 02:56:51 2021
Response via : Initial Calibration

