

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081621\
 Data File : VU044429.D
 Acq On : 16 Aug 2021 12:44
 Operator : SY/MD
 Sample : VSTD00113
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD001013

Quant Time: Aug 17 02:48:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 17 02:46:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	6232	0.500	ug/L	0.02
5) Chlorobenzene-d5	9.364	117	5811	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	2916	0.500	ug/L	-0.01
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	8267	1.018	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.680	65	8645	0.925	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.639	67	7423	1.039	ug/L	0.00
8) Toluene-d8	7.849	98	16428	1.005	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.682	84	6171	1.018	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.606	62	6107	0.979	ug/L	98
6) Trichloroethene	6.508	95	5952	1.019	ug/L	97
9) 1,2-Dibromoethane	8.869	107	4195	0.986	ug/L #	99
12) 1,2,3-Trichloropropane	10.751	75	4306	0.989	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.927	75	802	0.988	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081621\
Data File : VU044429.D
Acq On : 16 Aug 2021 12:44
Operator : SY/MD
Sample : VSTD00113
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD001013

Quant Time: Aug 17 02:48:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 17 02:46:12 2021
Response via : Initial Calibration

