

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081621\
 Data File : VU044431.D
 Acq On : 16 Aug 2021 13:47
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VICV015

Quant Time: Aug 17 03:02:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081621.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 17 02:56:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.225	114	5865	0.500	ug/L	0.02
5) Chlorobenzene-d5	9.364	117	5579	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.745	152	2789	0.500	ug/L	-0.01
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.599	65	3827	0.499	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	100.000%
4) 1,2-Dichloroethane-d4	5.680	65	4073	0.462	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
7) 1,2-Dichloropropane-d6	6.639	67	3493	0.509	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	102.000%
8) Toluene-d8	7.849	98	7725	0.492	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.682	84	2986	0.513	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.606	62	3150	0.535	ug/L	93
6) Trichloroethene	6.508	95	3001	0.535	ug/L	98
9) 1,2-Dibromoethane	8.870	107	2124	0.520	ug/L	100
12) 1,2,3-Trichloropropane	10.751	75	2253	0.541	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.927	75	414	0.533	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081621\
Data File : VU044431.D
Acq On : 16 Aug 2021 13:47
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VICV015

Quant Time: Aug 17 03:02:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081621.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 17 02:56:51 2021
Response via : Initial Calibration

