

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081624\
 Data File : VU060472.D
 Acq On : 16 Aug 2024 13:01
 Operator : MD/SY
 Sample : VSTD0.5
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Aug 17 04:56:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 04:56:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	5565	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	8477	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2599	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	4894	0.541	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	2515	0.524	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	2796	0.531	ug/L	0.00
8) Toluene-d8	7.887	98	5715	0.525	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	2471	0.535	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	4846	0.510	ug/L	100
6) Trichloroethene	6.544	95	2574	0.481	ug/L	100
9) 1,2-Dibromoethane	8.924	107	1735	0.517	ug/L	100
12) 1,2,3-Trichloropropane	10.814	75	1505	0.499	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.990	75	308	0.538	ug/L	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081624\
Data File : VU060472.D
Acq On : 16 Aug 2024 13:01
Operator : MD/SY
Sample : VSTD0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Aug 17 04:56:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 04:56:02 2024
Response via : Initial Calibration









#3
 Vinyl chloride
 Concen: 0.510 ug/L
 RT: 1.597 min Scan# 114
 Delta R.T. 0.000 min
 Lab File: VU060472.D
 Acq: 16 Aug 2024 13:01




Instrument :
 MSVOA_U
 ClientSampleId :


Tgt Ion: 62 Resp: 4846
 Ion Ratio Lower Upper
 62 100
 64 34.1 23.9 44.3



#6
 Trichloroethene
 Concen: 0.481 ug/L
 RT: 6.544 min Scan# 458
 Delta R.T. 0.000 min
 Lab File: VU060472.D
 Acq: 16 Aug 2024 13:01


Tgt Ion: 95 Resp: 2574
 Ion Ratio Lower Upper
 95 100
 97 69.3 48.5 90.1
 132 107.8 75.5 140.1
 130 109.0 76.3 141.7








#9
1,2-Dibromoethane
Concen: 0.517 ug/L
RT: 8.924 min Scan# 596
Delta R.T. 0.000 min
Lab File: VU060472.D
Acq: 16 Aug 2024 13:01

Instrument :
MSVOA_U
ClientSampleId :







Tgt Ion:107 Resp: 1735
Ion Ratio Lower Upper
107 100
109 95.0 66.5 123.5
188 12.0 9.6 14.4



#12
1,2,3-Trichloropropane
Concen: 0.499 ug/L
RT: 10.814 min Scan# 705
Delta R.T. 0.000 min
Lab File: VU060472.D
Acq: 16 Aug 2024 13:01


Tgt Ion: 75 Resp: 1505
Ion Ratio Lower Upper
75 100
110 40.5 32.4 48.6
77 32.5 26.0 39.0





#13
 1,2-Dibromo-3-chloropropane
 Concen: 0.538 ug/L
 RT: 12.990 min Scan# 832
 Delta R.T. 0.000 min
 Lab File: VU060472.D
 Acq: 16 Aug 2024 13:01

Instrument :
 MSVOA_U
 ClientSampleId :



Tgt Ion: 75 Resp: 308

Ion	Ratio	Lower	Upper
75	100		
157	92.9	74.3	111.5
155	82.7	66.2	99.2

