

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081624\
 Data File : VU060471.D
 Acq On : 16 Aug 2024 12:25
 Operator : MD/SY
 Sample : VSTD0.126
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.1026

Quant Time: Aug 17 04:56:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 04:56:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	5621	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	8320	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2526	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	832	0.091	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	437	0.090	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	460	0.089	ug/L	0.00
8) Toluene-d8	7.903	98	979	0.092	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.745	84	412	0.091	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	972	0.101	ug/L	92
6) Trichloroethene	6.544	95	567	0.108	ug/L	94
9) 1,2-Dibromoethane	8.924	107	344	0.104	ug/L #	92
12) 1,2,3-Trichloropropane	10.814	75	297	0.101	ug/L	96
13) 1,2-Dibromo-3-chloropr...	13.007	75	54	0.097	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081624\
Data File : VU060471.D
Acq On : 16 Aug 2024 12:25
Operator : MD/SY
Sample : VSTD0.126
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.1026

Quant Time: Aug 17 04:56:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 04:56:02 2024
Response via : Initial Calibration

