

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081624\
 Data File : VU060472.D
 Acq On : 16 Aug 2024 13:01
 Operator : MD/SY
 Sample : VSTD0.527
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5027

Quant Time: Aug 17 04:56:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 04:56:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	5565	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	8477	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2599	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	4894	0.541	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	2515	0.524	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	2796	0.531	ug/L	0.00
8) Toluene-d8	7.887	98	5715	0.525	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	2471	0.535	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	4846	0.510	ug/L	100
6) Trichloroethene	6.544	95	2574	0.481	ug/L	100
9) 1,2-Dibromoethane	8.924	107	1735	0.517	ug/L	100
12) 1,2,3-Trichloropropane	10.814	75	1505	0.499	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.990	75	308	0.538	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081624\
Data File : VU060472.D
Acq On : 16 Aug 2024 13:01
Operator : MD/SY
Sample : VSTD0.527
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5027

Quant Time: Aug 17 04:56:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 04:56:02 2024
Response via : Initial Calibration

