

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081720\
 Data File : VU039885.D
 Acq On : 17 Aug 2020 14:15
 Operator : SY/MD
 Sample : L3517-18 100X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F4M04

Quant Time: Aug 18 05:03:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR080320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 18 04:18:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	83988	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	83671	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	40371	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	16961	3.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.80%
7) Chloroethane-d5	1.92	69	20635	4.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.20%
11) 1,1-Dichloroethene-d2	2.58	63	27510	3.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.20%
20) 2-Butanone-d5	4.65	46	103346	51.81	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.62%
24) Chloroform-d	5.08	84	48477	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.60%
26) 1,2-Dichloroethane-d4	5.72	65	30244	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
32) Benzene-d6	5.75	84	87718	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
36) 1,2-Dichloropropane-d6	6.70	67	30733	4.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.20%
41) Toluene-d8	7.91	98	67752	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.80%
43) trans-1,3-Dichloropropene-	8.19	79	11412	4.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.40%
46) 2-Hexanone-d5	8.64	63	74683	52.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.60%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	29565	4.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	32366	4.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU081720\
Data File : VU039885.D
Acq On : 17 Aug 2020 14:15
Operator : SY/MD
Sample : L3517-18 100X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F4M04

Quant Time: Aug 18 05:03:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR080320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 18 04:18:42 2020
Response via : Initial Calibration

