

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081819\
 Data File : VU033832.D
 Acq On : 16 Aug 2019 13:45
 Operator : JC/SP
 Sample : K4369-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CD7

Quant Time: Aug 17 02:27:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM081619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 17 02:15:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	369814	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	372198	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	158140	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	143832	43.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.46%
7) Chloroethane-d5	1.67	69	128611	41.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.66%
11) 1,1-Dichloroethene-d2	2.26	63	184122	33.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.22%
21) 2-Butanone-d5	4.15	46	192389	110.26	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.26%
24) Chloroform-d	4.62	84	254553	47.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.06%
26) 1,2-Dichloroethane-d4	5.29	65	166858	49.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.62%
32) Benzene-d6	5.32	84	488574	47.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.24%
36) 1,2-Dichloropropane-d6	6.31	67	164002	48.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.04%
41) Toluene-d8	7.55	98	420444	43.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.26%
43) trans-1,3-Dichloropropene-	7.83	79	71201	45.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.34%
47) 2-Hexanone-d5	8.30	63	122436	98.35	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.35%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	222909	47.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.42%
64) 1,2-Dichlorobenzene-d4	11.84	152	169350	53.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.46%

Target Compounds					Ovalue
13) Acetone	2.32	43	5635	3.286 ug/L	87
25) Chloroform	4.65	83	53652	10.069 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU081819\
Data File : VU033832.D
Acq On : 16 Aug 2019 13:45
Operator : JC/SP
Sample : K4369-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CD7

Quant Time: Aug 17 02:27:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM081619WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 17 02:15:05 2019
Response via : Initial Calibration

