

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081821\
 Data File : VU044452.D
 Acq On : 18 Aug 2021 18:11
 Operator : SY/MD
 Sample : M3474-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD2

Quant Time: Aug 19 04:05:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Aug 19 03:54:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	115384	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	117957	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	51684	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	44613	4.781	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.600%
7) Chloroethane-d5	1.916	69	40817	5.033	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.600%
11) 1,1-Dichloroethene-d2	2.571	65	19598	4.937	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.800%
20) 2-Butanone-d5	4.639	46	132517	51.346	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.700%
24) Chloroform-d	5.070	84	80047	4.851	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
26) 1,2-Dichloroethane-d4	5.710	65	46551	5.173	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.400%
32) Benzene-d6	5.735	84	164457	4.951	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	6.697	67	54263	5.107	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.200%
41) Toluene-d8	7.906	98	139740	4.744	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.800%
43) trans-1,3-Dichloroprop...	8.185	79	17648	4.624	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.400%
46) 2-Hexanone-d5	8.642	63	105681	51.210	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.420%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	42700	4.771	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%
66) 1,2-Dichlorobenzene-d4	12.198	152	47472	5.436	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081821\
Data File : VU044452.D
Acq On : 18 Aug 2021 18:11
Operator : SY/MD
Sample : M3474-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD2

Quant Time: Aug 19 04:05:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Aug 19 03:54:51 2021
Response via : Initial Calibration

