

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081919\
 Data File : VU033837.D
 Acq On : 19 Aug 2019 09:22
 Operator : JC/SP
 Sample : VU0819WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK56

Quant Time: Aug 19 15:36:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR081419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 01:44:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	178239	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	184132	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.47	152	82070	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	68367	3.39	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	67.80%
7) Chloroethane-d5	1.67	69	63032	4.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.00%
11) 1,1-Dichloroethene-d2	2.26	65	33331	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	4.19	46	208103	53.39	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.78%
24) Chloroform-d	4.62	84	121719	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
26) 1,2-Dichloroethane-d4	5.29	65	69796	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.60%
32) Benzene-d6	5.31	84	226880	4.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.20%
36) 1,2-Dichloropropane-d6	6.31	67	76576	4.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.40%
41) Toluene-d8	7.55	98	183922	3.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.80%
43) trans-1,3-Dichloropropene-	7.84	79	29880	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.30	63	141703	50.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.18%
56) 1,1,2,2-Tetrachloroethane-	10.42	84	70438	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.80%
66) 1,2-Dichlorobenzene-d4	11.85	152	76388	4.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.80%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081919\
 Data File : VU033837.D
 Acq On : 19 Aug 2019 09:22
 Operator : JC/SP
 Sample : VU0819WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampled :
 VBLK56

Quant Time: Aug 19 15:36:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUTR081419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 01:44:12 2019
 Response via : Initial Calibration

