

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081921\
 Data File : VU044469.D
 Acq On : 19 Aug 2021 17:44
 Operator : SY/MD
 Sample : M3474-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AF9

Quant Time: Aug 20 05:15:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 20 05:12:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.257	114	113356	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	114870	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	51446	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	42886	4.678	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.600%
7) Chloroethane-d5	1.916	69	40344	5.063	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.200%
11) 1,1-Dichloroethene-d2	2.572	65	19022	4.878	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.600%
20) 2-Butanone-d5	4.639	46	135428	53.413	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.820%
24) Chloroform-d	5.070	84	79531	4.906	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	5.710	65	46118	5.217	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.736	84	163425	5.052	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
36) 1,2-Dichloropropane-d6	6.697	67	53029	5.125	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.400%
41) Toluene-d8	7.903	98	137065	4.779	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
43) trans-1,3-Dichloroprop...	8.186	79	17492	4.707	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.200%
46) 2-Hexanone-d5	8.646	63	108002	53.741	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.480%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	43465	4.987	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.800%
66) 1,2-Dichlorobenzene-d4	12.198	152	46841	5.389	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081921\
Data File : VU044469.D
Acq On : 19 Aug 2021 17:44
Operator : SY/MD
Sample : M3474-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AF9

Quant Time: Aug 20 05:15:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 20 05:12:49 2021
Response via : Initial Calibration

