

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081921\
 Data File : VU044470.D
 Acq On : 19 Aug 2021 18:09
 Operator : SY/MD
 Sample : M3474-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD6

Manual Integrations
 APPROVED

MMDadoda
 8/20/2021 12:49:25 PM

Quant Time: Aug 20 05:15:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 20 05:12:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	111387	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	112985	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	50814	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	42129	4.676	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.600%
7) Chloroethane-d5	1.916	69	38030	4.857	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.200%
11) 1,1-Dichloroethene-d2	2.572	65	19331	5.044	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.800%
20) 2-Butanone-d5	4.642	46	132375m	53.132	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.260%
24) Chloroform-d	5.070	84	79214	4.973	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
26) 1,2-Dichloroethane-d4	5.710	65	44072	5.073	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.400%
32) Benzene-d6	5.735	84	157344	4.945	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	6.697	67	51833	5.092	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.800%
41) Toluene-d8	7.903	98	132674	4.703	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	8.186	79	17187	4.702	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.000%
46) 2-Hexanone-d5	8.645	63	105361	53.302	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.600%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	41982	4.897	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	12.198	152	45828	5.338	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081921\
Data File : VU044470.D
Acq On : 19 Aug 2021 18:09
Operator : SY/MD
Sample : M3474-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD6

Manual Integrations
APPROVED

MMDadoda
8/20/2021 12:49:25 PM

Quant Time: Aug 20 05:15:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 20 05:12:49 2021
Response via : Initial Calibration

