

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081921\
 Data File : VU044472.D
 Acq On : 19 Aug 2021 18:57
 Operator : SY/MD
 Sample : M3474-22
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AG2

Manual Integrations
 APPROVED

MMDadoda
 8/20/2021 12:49:32 PM

Quant Time: Aug 20 05:15:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 20 05:12:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	116353	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	117476	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	53056	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	41427	4.402	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.000%	
7) Chloroethane-d5	1.916	69	38215	4.673	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.400%	
11) 1,1-Dichloroethene-d2	2.572	65	18544	4.632	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	92.600%	
20) 2-Butanone-d5	4.642	46	130425m	50.115	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.220%	
24) Chloroform-d	5.070	84	76964	4.625	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.600%	
26) 1,2-Dichloroethane-d4	5.710	65	44367	4.889	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.800%	
32) Benzene-d6	5.735	84	158543	4.793	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	6.697	67	51386	4.856	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.200%	
41) Toluene-d8	7.903	98	132333	4.511	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.200%	
43) trans-1,3-Dichloroprop...	8.186	79	17316	4.556	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.200%	
46) 2-Hexanone-d5	8.645	63	105997	51.574	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	103.140%	
56) 1,1,2,2-Tetrachloroeth...	10.764	84	42237	4.739	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	94.800%	
66) 1,2-Dichlorobenzene-d4	12.198	152	45983	5.130	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081921\
Data File : VU044472.D
Acq On : 19 Aug 2021 18:57
Operator : SY/MD
Sample : M3474-22
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AG2

Manual Integrations
APPROVED

MMDadoda
8/20/2021 12:49:32 PM

Quant Time: Aug 20 05:15:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 20 05:12:49 2021
Response via : Initial Calibration

