

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082018\
Data File : VU026206.D
Acq On : 20 Aug 2018 14:09
Operator : MD/SY
Sample : VU0820WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0820WBL01

Quant Time: Aug 21 08:24:47 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U082018DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Aug 20 13:35:41 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.75	96	33416	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.31	95	10047	0.84	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	11444	0.82	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.00%	

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082018\
Data File : VU026206.D
Acq On : 20 Aug 2018 14:09
Operator : MD/SY
Sample : VU0820WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0820WBL01

Quant Time: Aug 21 08:24:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U082018DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Aug 20 13:35:41 2018
Response via : Initial Calibration

