

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082018\
 Data File : VU026215.D
 Acq On : 20 Aug 2018 17:58
 Operator : MD/SY
 Sample : J4045-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PT-ADD-WS

Quant Time: Aug 21 08:46:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U082018DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Mon Aug 20 13:35:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.75	96	34973	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.31	95	9126	0.73	ug/l	0.00
Spiked Amount 1.000			Recovery	=	73.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	11749	0.80	ug/l	0.00
Spiked Amount 1.000			Recovery	=	80.00%	

Target Compounds

						Ovalue
8) 1,1,2-Trichloro-1,2,2-trif	2.29	101	137800	14.49	ug/l	99
24) tert-Butyl Alcohol	2.84	59	12599	19.79	ug/l #	87
25) Methyl tert-Butyl Ether	3.00	73	1012676	47.82	ug/l	98
31) Isopropyl Ether	3.58	45	1335847	47.16	ug/l	96
32) Ethyl-t-butyl ether	4.08	59	192197	7.75	ug/l	93
33) Tert-Amyl methyl ether	5.59	73	1156044	54.54	ug/l	98
73) n-propylbenzene	10.59	120	87255	7.01	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082018\
Data File : VU026215.D
Acq On : 20 Aug 2018 17:58
Operator : MD/SY
Sample : J4045-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Aug 21 08:46:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U082018DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Aug 20 13:35:41 2018
Response via : Initial Calibration

