

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU082019\
 Data File : VU033853.D
 Acq On : 20 Aug 2019 14:55
 Operator : JC/SP
 Sample : K4373-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AF4

Quant Time: Aug 21 07:39:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR082019WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 21 05:39:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	204371	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	204586	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	98262	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	71068	4.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.60%
7) Chloroethane-d5	1.67	69	64929	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.60%
11) 1,1-Dichloroethene-d2	2.26	63	119863	3.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.60%
20) 2-Butanone-d5	4.16	46	132206	38.29	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.58%
24) Chloroform-d	4.62	84	125314	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.40%
26) 1,2-Dichloroethane-d4	5.29	65	65030	4.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.00%
32) Benzene-d6	5.31	84	249978	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	6.31	67	78664	4.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.20%
41) Toluene-d8	7.55	98	212111	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
43) trans-1,3-Dichloropropene-	7.83	79	27462	4.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.60%
46) 2-Hexanone-d5	8.30	63	97641	42.67	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.34%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	63548	4.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.40%
64) 1,2-Dichlorobenzene-d4	11.84	152	88928	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%


Target Compounds					Qvalue
13) Acetone	2.31	43	12132	4.596 ug/L	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU082019\
Data File : VU033853.D
Acq On : 20 Aug 2019 14:55
Operator : JC/SP
Sample : K4373-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AF4

Quant Time: Aug 21 07:39:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR082019WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 21 05:39:59 2019
Response via : Initial Calibration





#13

Acetone

Concen: 4.596 ug/L

RT: 2.31 min Scan# 378

Delta R.T. -0.00 min

Lab File: VU033853.D

Acq: 20 Aug 2019 14:55

Instrument :
MSVOA_U
ClientSampleId :
C0AF4

Tgt Ion: 43 Resp: 12132

Ion Ratio Lower Upper

43 100

58 37.6 0.0 65.4

