

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082024\
 Data File : VU060488.D
 Acq On : 20 Aug 2024 14:20
 Operator : MD/SY
 Sample : P3627-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Aug 21 02:31:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 21 02:30:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	78894	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	86896	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	36490	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	25007	4.038	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.800%
7) Chloroethane-d5	1.911	69	24234	4.340	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.800%
11) 1,1-Dichloroethene-d2	2.560	65	11513	4.368	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.400%
20) 2-Butanone-d5	4.634	46	62519	47.570	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.140%
24) Chloroform-d	5.055	84	57604	4.698	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
26) 1,2-Dichloroethane-d4	5.695	65	29039	5.107	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
32) Benzene-d6	5.721	84	106396	4.060	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%
36) 1,2-Dichloropropane-d6	6.682	67	35733	4.396	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.000%
41) Toluene-d8	7.891	98	87087	3.722	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.400%
43) trans-1,3-Dichloroprop...	8.177	79	11293	4.297	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.631	63	45790	39.264	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.520%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	28863	4.327	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	31831	4.735	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.800%
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	1905	0.258	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082024\
Data File : VU060488.D
Acq On : 20 Aug 2024 14:20
Operator : MD/SY
Sample : P3627-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Aug 21 02:31:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 21 02:30:39 2024
Response via : Initial Calibration

