

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082118\
 Data File : VU026222.D
 Acq On : 21 Aug 2018 10:50
 Operator : MD/SY
 Sample : VU0821WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK41

Quant Time: Aug 22 00:28:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR081718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 22 00:24:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	146388	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	146851	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	67725	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	37476	4.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.00%
7) Chloroethane-d5	1.68	69	31689	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.28	63	58302	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	4.19	46	116307	52.69	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.38%
24) Chloroform-d	4.65	84	78094	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.40%
26) 1,2-Dichloroethane-d4	5.31	65	47627	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	5.34	84	166103	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	6.34	67	52824	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.40%
41) Toluene-d8	7.57	98	144935	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
43) trans-1,3-Dichloropropene-	7.85	79	21641	4.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.60%
46) 2-Hexanone-d5	8.32	63	92165	50.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.76%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	42099	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	60383	5.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082118\
Data File : VU026222.D
Acq On : 21 Aug 2018 10:50
Operator : MD/SY
Sample : VU0821WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK41

Quant Time: Aug 22 00:28:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR081718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 22 00:24:50 2018
Response via : Initial Calibration

