

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082118\
 Data File : VU026227.D
 Acq On : 21 Aug 2018 12:59
 Operator : MD/SY
 Sample : J4515-03DL 40X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0326DL

Quant Time: Aug 22 00:38:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR081718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Aug 22 00:24:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	129037	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	131516	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	60429	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	35107	5.15	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.00%
7) Chloroethane-d5	1.68	69	29139	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.20%
11) 1,1-Dichloroethene-d2	2.28	63	68877	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.40%
20) 2-Butanone-d5	4.19	46	110955	57.02	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.04%
24) Chloroform-d	4.65	84	84395	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.60%
26) 1,2-Dichloroethane-d4	5.31	65	46007	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.20%
32) Benzene-d6	5.34	84	157717	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	6.34	67	50909	5.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.80%
41) Toluene-d8	7.57	98	135853	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
43) trans-1,3-Dichloropropene-	7.85	79	19699	5.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.20%
46) 2-Hexanone-d5	8.32	63	87130	53.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.36%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	40937	5.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	57177	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%


Target Compounds						Ovalue
12) 1,1-Dichloroethene	2.29	96	25654	3.42	ug/L	89
19) 1,1-Dichloroethane	3.45	63	3355	0.22	ug/L	95
29) 1,1,1-Trichloroethane	4.92	97	165045	9.69	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082118\
Data File : VU026227.D
Acq On : 21 Aug 2018 12:59
Operator : MD/SY
Sample : J4515-03DL 40X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0326DL

Quant Time: Aug 22 00:38:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR081718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Aug 22 00:24:50 2018
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 3.42 ug/L

RT: 2.29 min Scan# 528

Delta R.T. 0.00 min

Lab File: VU026227.D

Acq: 21 Aug 2018 12:59

Instrument :

MSVOA_U

ClientSampleId :

C0326DL





Tot Ion: 96 Resp: 25654

Ion Ratio Lower Upper

96 100

61 184.5 125.3 232.7

63 153.2 90.9 168.9



#19

1,1-Dichloroethane

Concen: 0.22 ug/L

RT: 3.45 min Scan# 889

Delta R.T. 0.00 min

Lab File: VU026227.D

Acq: 21 Aug 2018 12:59




Tgt Ion: 63 Resp: 3355


Ion Ratio Lower Upper

63 100

65 35.2 23.2 43.0

83 11.0 10.1 18.9





#29

1,1,1-Trichloroethane

Concen: 9.69 ug/L

RT: 4.92 min Scan# 1346

Delta R.T. 0.00 min

Lab File: VU026227.D

Acq: 21 Aug 2018 12:59

Instrument :

MSVOA_U

ClientSampleId :

C0326DL

Tot Ion: 97 Resp: 165045

Ion Ratio Lower Upper

97 100

99 63.7 52.3 78.5

61 42.9 35.0 52.6

