

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU082218\
 Data File : VU026271.D
 Acq On : 23 Aug 2018 00:33
 Operator : MD/SY
 Sample : J4599-01 40X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETBN8

Quant Time: Aug 23 07:47:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 23 07:46:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	283839	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	282343	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	132039	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	71466	41.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.56%
7) Chloroethane-d5	1.68	69	62191	44.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.10%
11) 1,1-Dichloroethene-d2	2.27	63	110802	32.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.12%
21) 2-Butanone-d5	4.18	46	134936	97.66	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.66%
24) Chloroform-d	4.65	84	174172	45.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.28%
26) 1,2-Dichloroethane-d4	5.31	65	116037	46.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.42%
32) Benzene-d6	5.35	84	334003	44.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.44%
36) 1,2-Dichloropropane-d6	6.33	67	110881	45.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.44%
41) Toluene-d8	7.57	98	299704	41.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.02%
43) trans-1,3-Dichloropropene-	7.85	79	52489	41.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.78%
47) 2-Hexanone-d5	8.31	63	92772	89.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.50%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	155447	43.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.62%
64) 1,2-Dichlorobenzene-d4	11.86	152	133964	47.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.62%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU082218\
Data File : VU026271.D
Acq On : 23 Aug 2018 00:33
Operator : MD/SY
Sample : J4599-01 40X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETBN8

Quant Time: Aug 23 07:47:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 23 07:46:41 2018
Response via : Initial Calibration

