

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082222\
 Data File : VU050393.D
 Acq On : 22 Aug 2022 11:24
 Operator : SY/MD
 Sample : VU0822WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0822WBL01

Quant Time: Aug 23 08:20:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Aug 23 08:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.112	96	31503	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	8078	0.753	ug/l	0.00
Spiked Amount 1.000			Recovery	=	75.000%	
68) 1,2-Dichlorobenzene-d4	12.198	152	9837	0.776	ug/l	0.00
Spiked Amount 1.000			Recovery	=	78.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082222\
Data File : VU050393.D
Acq On : 22 Aug 2022 11:24
Operator : SY/MD
Sample : VU0822WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0822WBL01

Quant Time: Aug 23 08:20:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Aug 23 08:18:20 2022
Response via : Initial Calibration

