

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082222\
 Data File : VU050399.D
 Acq On : 22 Aug 2022 14:32
 Operator : SY/MD
 Sample : N3734-14DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WS-NAPHTHALENEDL

Quant Time: Aug 23 08:27:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Aug 23 08:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.108	96	31362	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.639	95	8479	0.794	ug/l	0.00
Spiked Amount 1.000			Recovery	=	79.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	10076	0.798	ug/l	0.00
Spiked Amount 1.000			Recovery	=	80.000%	
Target Compounds						
14) Carbon Disulfide	2.790	76	65073	2.274	ug/l	98
45) 4-Methyl-2-Pentanone	7.819	43	13613	2.421	ug/l	94
89) Naphthalene	14.092	128	87651	3.571	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082222\
Data File : VU050399.D
Acq On : 22 Aug 2022 14:32
Operator : SY/MD
Sample : N3734-14DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WS-NAPHTHALENEDL

Quant Time: Aug 23 08:27:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Aug 23 08:18:20 2022
Response via : Initial Calibration

