

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU082318\
 Data File : VU026298.D
 Acq On : 23 Aug 2018 15:04
 Operator : MD/SY
 Sample : J4598-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETBN7

Quant Time: Aug 24 01:45:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 24 01:14:40 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	258535	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	256704	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	121097	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	81468	51.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.32%
7) Chloroethane-d5	1.68	69	67825	52.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.50%
11) 1,1-Dichloroethene-d2	2.27	63	125084	40.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.72%
21) 2-Butanone-d5	4.18	46	132377	105.18	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.18%
24) Chloroform-d	4.65	84	181719	51.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.40%
26) 1,2-Dichloroethane-d4	5.31	65	122388	53.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.02%
32) Benzene-d6	5.35	84	356164	51.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.72%
36) 1,2-Dichloropropane-d6	6.33	67	115169	52.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.46%
41) Toluene-d8	7.57	98	322970	49.20	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.40%
43) trans-1,3-Dichloropropene-	7.85	79	54775	48.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.16%
47) 2-Hexanone-d5	8.31	63	94685	100.47	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.47%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	159475	48.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.74%
64) 1,2-Dichlorobenzene-d4	11.86	152	137347	53.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.90%

Target Compounds					Ovalue
13) Acetone	2.32	43	5266	5.41 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU082318\
Data File : VU026298.D
Acq On : 23 Aug 2018 15:04
Operator : MD/SY
Sample : J4598-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETBN7

Quant Time: Aug 24 01:45:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM082218WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 24 01:14:40 2018
Response via : Initial Calibration

