

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082321\
 Data File : VU044506.D
 Acq On : 23 Aug 2021 17:30
 Operator : SY/MD
 Sample : M3517-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AG3

Manual Integrations
 APPROVED

SAM
 8/24/2021 11:21:07 AM

Quant Time: Aug 24 06:17:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 24 06:16:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	117536	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	120203	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	53796	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	37718	3.968	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.400%
7) Chloroethane-d5	1.919	69	36590	4.429	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.600%
11) 1,1-Dichloroethene-d2	2.571	65	17683	4.373	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.400%
20) 2-Butanone-d5	4.639	46	130799m	49.753	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.500%
24) Chloroform-d	5.073	84	76063	4.525	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	5.710	65	42683	4.657	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
32) Benzene-d6	5.735	84	155112	4.583	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.697	67	50768	4.688	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.800%
41) Toluene-d8	7.906	98	130186	4.337	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
43) trans-1,3-Dichloroprop...	8.185	79	16884	4.342	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.800%
46) 2-Hexanone-d5	8.645	63	107501	51.119	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.240%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	42044	4.610	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	12.198	152	44788	4.928	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082321\
Data File : VU044506.D
Acq On : 23 Aug 2021 17:30
Operator : SY/MD
Sample : M3517-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AG3

Manual Integrations
APPROVED

SAM
8/24/2021 11:21:07 AM

Quant Time: Aug 24 06:17:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 24 06:16:49 2021
Response via : Initial Calibration

