

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082321\
 Data File : VU044494.D
 Acq On : 23 Aug 2021 11:51
 Operator : SY/MD
 Sample : M3515-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4555

Quant Time: Aug 24 03:58:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 24 03:56:17 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	127740	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	129754	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	58822	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	37009	3.582	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.600%
7) Chloroethane-d5	1.916	69	36070	4.017	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.400%
11) 1,1-Dichloroethene-d2	2.572	65	16676	3.794	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.800%
20) 2-Butanone-d5	4.639	46	129462	45.310	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.620%
24) Chloroform-d	5.073	84	76757	4.202	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
26) 1,2-Dichloroethane-d4	5.710	65	43620	4.379	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
32) Benzene-d6	5.735	84	153645	4.205	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.697	67	50792	4.345	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.000%
41) Toluene-d8	7.906	98	129532	3.998	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.000%
43) trans-1,3-Dichloroprop...	8.186	79	17185	4.094	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.800%
46) 2-Hexanone-d5	8.642	63	105412	46.436	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.880%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	41713	4.237	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.800%
66) 1,2-Dichlorobenzene-d4	12.198	152	44375	4.465	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.400%
Target Compounds						
2) Dichlorodifluoromethane	1.395	85	1267	0.096	ug/L	93
3) Chloromethane	1.523	50	13378	0.777	ug/L	96
17) Methyl tert-butyl Ether	3.366	73	5551	0.219	ug/L	96
25) Chloroform	5.099	83	3730	0.183	ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU082321\
Data File : VU044494.D
Acq On : 23 Aug 2021 11:51
Operator : SY/MD
Sample : M3515-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4555

Quant Time: Aug 24 03:58:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR081821WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 24 03:56:17 2021
Response via : Initial Calibration









#2
Dichlorodifluoromethane
Concen: 0.096 ug/L
RT: 1.395 min Scan# 17
Delta R.T. 0.007 min
Lab File: VU044494.D
Acq: 23 Aug 2021 11:51




Instrument :
MSVOA_U
ClientSampled :
H4555


Tgt Ion: 85 Resp: 1267
Ion Ratio Lower Upper
85 100
87 36.2 25.7 38.5



#3
Chloromethane
Concen: 0.777 ug/L
RT: 1.523 min Scan# 57
Delta R.T. 0.000 min
Lab File: VU044494.D
Acq: 23 Aug 2021 11:51

Tgt Ion: 50 Resp: 13378
Ion Ratio Lower Upper
50 100
52 30.3 22.8 42.4









#17
Methyl tert-butyl Ether
Concen: 0.219 ug/L
RT: 3.366 min Scan# 630
Delta R.T. -0.006 min
Lab File: VU044494.D
Acq: 23 Aug 2021 11:51

Instrument :
MSVOA_U
ClientSampled :
H4555

Tgt Ion: 73	Resp: 5551
Ion Ratio	Lower Upper
73 100	
43 22.5	16.2 24.4
57 20.5	17.8 26.8



#25
Chloroform
Concen: 0.183 ug/L
RT: 5.099 min Scan# 1169
Delta R.T. 0.000 min
Lab File: VU044494.D
Acq: 23 Aug 2021 11:51

Tgt Ion: 83	Resp: 3730
Ion Ratio	Lower Upper
83 100	
85 59.4	45.3 84.1

